

Certificate of Calibration - Wind Monitoring Station

Description: Yau Lai Estate, Bik Lai House

Manufacturer: <u>Davis Instruments</u>

Model No.: <u>Davis7440</u>

Serial No.: MC01010A44

Equipment No.: <u>SA-03-04</u>

Date of Calibration <u>17-Feb-2025</u>

Next Due Date <u>17-Aug-2025</u>

1. Performance check of Wind Speed

Wind Sp	peed, m/s	Difference D (m/s)
Wind Speed Reading (V1)	Anemometer Value (V2)	D = V1 - V2
0.0	0.0	0.0
1.5	1.4	0.1
2.5	2.4	0.1
4.0	3.8	0.2

2. Performance check of Wind Direction

Wind Di	rection (°)	Difference D (°)
Wind Direction Reading (W1)	Marine Compass Value (W2)	D = W1 - W2
0	0	0.0
90	90	0.0
180	180	0.0
270	270	0.0

Test Specification:

- 1. Performance Wind Speed Test The wind meter was on-site calibrated against the anemometer
- 2. Performance Wind Direction Test The wind meter was on-site calibrated against the marine compass at four direction

Calibrated by:

Wong Shing Kwai

Approved by:

Henry/Leung



Certificate of Calibration - Wind Monitoring Station

Description: Yau Lai Estate, Bik Lai House

Manufacturer: <u>Davis Instruments</u>

Model No.: <u>Davis7440</u>

Serial No.: MC01010A44

Equipment No.: SA-03-04

Date of Calibration <u>17-Aug-2025</u>

Next Due Date <u>17-Feb-2026</u>

1. Performance check of Wind Speed

Wind Sp	peed, m/s	Difference D (m/s)
Wind Speed Reading (V1)	Anemometer Value (V2)	D = V1 - V2
0.0	0.0	0.0
1.5	1.5	0.0
2.5	2.4	0.1
4.0	3.9	0.1

2. Performance check of Wind Direction

Wind Di	rection (°)	Difference D (°)
Wind Direction Reading (W1)	Marine Compass Value (W2)	D = W1 - W2
0	0	0.0
90	90	0.0
180	180	0.0
270	270	0.0

Test Specification:

- 1. Performance Wind Speed Test The wind meter was on-site calibrated against the anemometer
- 2. Performance Wind Direction Test The wind meter was on-site calibrated against the marine compass at four direction

Calibrated by:

Wong Shing Kwai

Approved by:

Henry Leung





RECALIBRATION DUE DATE:

January 7, 2026

Certificate of Calibration

Calibration Certification Information

Cal. Date: January 7, 2025 Rootsmeter S/N: 438320 Ta: 293 °K

Operator: Jim Tisch Pa: 759.0 mm Hg

Calibration Model #: TE-5025A Calibrator S/N: 3864

Run	Vol. Init (m3)	Vol. Final (m3)	ΔVol. (m3)	ΔTime (min)	ΔP (mm Hg)	ΔH (in H2O)
1	1	2	1	1.4590	3.2	2.00
2	3	4	1	1.0360	6.4	4.00
3	5	6	1	0.9160	8.0	5.00
4	7	8	1	0.8800	8.8	5.50
5	9	10	1	0.7270	12.7	8.00

Data Tabulation							
Vstd	Qstd	$\sqrt{\Delta H \left(\frac{Pa}{Pstd}\right) \left(\frac{Tstd}{Ta}\right)}$		Qa	√∆H(Ta/Pa)		
(m3)	(x-axis)	(y-axis)	Va	(x-axis)	(y-axis)		
1.0114	0.6932	1.4252	0.9958	0.6825	0.8787		
1.0071	0.9721	2.0156	0.9916	0.9571	1.2427		
1.0050	1.0971	2.2535	0.9895	1.0802	1.3893		
1.0039	1.1408	2.3635	0.9884	1.1232	1.4572		
0.9987	1.3737	2.8505	0.9833	1.3525	1.7574		
	m=	2.08969		m=	1.30853		
QSTD	b=	-0.02374	QA [b=	-0.01464		
,	r=	0.99985	,	r=	0.99985		

Calculations						
Vstd=	ΔVol((Pa-ΔP)/Pstd)(Tstd/Ta)	Va=	ΔVol((Pa-ΔP)/Pa)			
Qstd=	Vstd/∆Time	Qa=	Va/ΔTime			
	For subsequent flow rate calculations:					
Qstd=	$1/m \left(\sqrt{\Delta H \left(\frac{Pa}{Pstd} \right) \left(\frac{Tstd}{Ta} \right)} \right) - b $	Qa=	$1/m\left(\left(\sqrt{\Delta H\left(Ta/Pa\right)}\right)-b\right)$			

Standard Conditions					
Tstd:	298.15 ° _K				
Pstd:	760 mm Hg				
	Key				
	ΔH: calibrator manometer reading (in H2O)				
ΔP: rootsme	ΔP: rootsmeter manometer reading (mm Hg)				
	Ta: actual absolute temperature (°K)				
Pa: actual barometric pressure (mm Hg)					
b: intercept					
m: slope					

RECALIBRATION

US EPA recommends annual recalibration per 1998 40 Code of Federal Regulations Part 50 to 51, Appendix B to Part 50, Reference Method for the Determination of Suspended Particulate Matter in the Atmosphere, 9.2.17, page 30

FAX: (513)467-9009

High-Volume TSP Sampler 5-POINT CALIBRATION DATA SHEET



						File No. N	MA20003/18/033
Project No.	CKL 1 - Flat 121	Cha Kwo Ling	Village				
Date:	4-Ju	1-25	Next Due Date: 4-Sep-25		Operator:	SK	
Equipment No.:	A-01			: TE 5170			
			Ambient (Condition			
Temperatur	re, Ta (K)	303.8	Pressure, Pa	(mmHg)		754.1	
		0.	rifice Transfer Sta	andard Inform	ection		
Serial	No	3864	Slope, mc	0.05914	Intercept	t be	-0.02377
Last Calibra		7-Jan-25			$c = [\Delta H \times (Pa/76)]$		
Next Calibra		7-Jan-26			x (Pa/760) x (298)		
				<u> </u>		/1 /	
			Calibration of	TSP Sampler			
Calibration		Or	fice			HVS	
Point	ΔH (orifice), in. of water	[ΔH x (Pa/76	0) x (298/Ta)] ^{1/2}	Qstd (CFM) X - axis	ΔW (HVS), in. of water		x (298/Ta)] ^{1/2} Y-
1	13.2	3.58		61.01	8.8		2.93
2	10.4		3.18	54.20	7.3		2.67
3	8.3	2.84		48.46	5.1		2.23
4	6.2	2	2.46	41.94	3.4	1.82	
5	3.3		1.79	30.71	1.4		1.17
By Linear Regressions Slope, mw = Correlation of		_	9 970	Intercept, bw =	-0.667	72	
*If Correlation C	Coefficient < 0.99	0, check and rec	alibrate.				
			Set Point C	Calculation			
From the TSP Fi	eld Calibration C	urve, take Qstd	= 43 CFM				
From the Regress	sion Equation, th	e "Y" value acco	ording to				
			$Qstd + bw = [\Delta W]$	- (D - /7(0) - (3	100 /T - \11/2		
		mw x C	$\mathbf{vsta} + \mathbf{bw} = \mathbf{v}$	X (Pa//60) X (2	(98/1a)]		
Therefore, Se	et Point; W = (m	w x Qstd + bw)	² x (760 / Pa) x (⁷	Γa / 298) =	3.73		
Remarks:							
Conducted by:	Wong Sh	ing Kwai	Signature:	<u> </u>	<u></u>	Date:	4-Jul-25
Checked by:	Henry	Leung	Signature:	- lem	, Oron	Date:	4-Jul-25

High-Volume TSP Sampler 5-POINT CALIBRATION DATA SHEET



File No. MA20003/55/033 CKL 2 - Flat 103 Cha Kwo Ling Village Project No. 4-Jul-25 Next Due Date: 4-Sep-25 Date: Operator: SK Equipment No.: A-01-55 Model No.: TE 5170 Serial No. 1956 **Ambient Condition** Temperature, Ta (K) 303.8 Pressure, Pa (mmHg) 754.1 **Orifice Transfer Standard Information** 0.05914 Intercept, bc 3864 Slope, mc -0.02377 Serial No. mc x Qstd + bc = $[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$ 7-Jan-25 Last Calibration Date: Qstd = $\{ [\Delta H \times (Pa/760) \times (298/Ta)]^{1/2} -bc \} / mc$ Next Calibration Date: 7-Jan-26 **Calibration of TSP Sampler** Orfice Calibration $[\Delta W \times (Pa/760) \times (298/Ta)]^{1/2}$ ΔH (orifice), Ostd (CFM) ΔW (HVS), in. $[\Delta H \ x \ (Pa/760) \ x \ (298/Ta)]^{1/2}$ Point in. of water X - axis of water Y-axis 2.98 1 13.4 3.61 61.47 9.1 11.2 7.2 2 3.30 56.23 2.65 9.3 3.01 51.27 2.31 4 5.1 2.23 38.07 2.6 1.59 5 3.7 2.0 1.90 32.49 1.40 By Linear Regression of Y on X Slope , mw = 0.0551 Intercept, bw = -0.4575 Correlation coefficient* = 0.9969 *If Correlation Coefficient < 0.990, check and recalibrate. **Set Point Calculation** From the TSP Field Calibration Curve, take Qstd = 43 CFM From the Regression Equation, the "Y" value according to mw x Qstd + bw = $[\Delta W \times (Pa/760) \times (298/Ta)]^{1/2}$ Therefore, Set Point; W = $(\text{mw x Qstd} + \text{bw})^2 \times (760 / \text{Pa}) \times (\text{Ta} / 298) =$ 3.76 Remarks: Conducted by: Wong Shing Kwai Checked by: Henry Leung

High-Volume TSP Sampler

5-POINT CALIBRATION DATA SHEET



File No. MA20003/04/0031

Project No.	KER 1 - Future	Residential Dev	elopment at Kerry	Godown			
Date:	11-J	ul-25	Next Due Date:	11-	Sep-25	Operator:	SK
Equipment No.:	A-0	01-04	Model No.:	TE	E 5170	Serial No.	10595
			Ambient C	Condition			
Temperatu	re, Ta (K)	300.9	Pressure, Pa	(mmHg)		749.2	
			rifice Transfer Sta				
1/2						-0.02377	
		7-Jan-25			$c = [\Delta H \ X \ (Pa / 760)]$ (Pa / 760) $X \ (298 / 760)$		
Next Calibr	ation Date:	7-Jan-26		Qsta = { ΔH x	(Pa//00) X (298/	1 a) -bc} / mc	<u>:</u>
		<u>.</u>	Calibration of	TSP Sampler			
G-171		0	rfice			HVS	
Calibration Point	ΔH (orifice), in. of water	[ΔH x (Pa/7	60) x (298/Ta)] ^{1/2}	Qstd (CFM) X - axis	ΔW (HVS), in. of water		0) x (298/Ta)] ^{1/2} -axis
1	12.6		3.51	59.71	8.3	2	2.85
2	10.5		3.20	54.54	7.1	2	2.63
3	8.5		2.88	49.11	5.2	2	2.25
4	5.3		2.27	38.87	2.8	1	1.65
5	3.6		1.87	32.10	2.1	1	1.43
By Linear Regr Slope, mw = Correlation *If Correlation C	0.0538 coefficient* =).9951	Intercept, bw =	-0.355	56	
			Set Point C	alculation			
From the TSP Fi From the Regres	sion Equation, th	ne "Y" value acc			98/Ta)] ^{1/2} 3.92		
			, (, , (
Remarks:							
				h	<u> </u>		
Conducted by:	Wong Sl	ning Kwai	_ Signature:		<u> </u>	Date:	11-Jul-25
Checked by:	Henry	Leung	Signature:	\-len	, ohon	Date:	11-Jul-25

High-Volume TSP Sampler

5-POINT CALIBRATION DATA SHEET



File No. MA20003/44/0030

Project No.	KTD1 - Centre	of Excellence in	Paediatrics (Childr	en's Hospital)		_	
Date:	11-J	ul-25	Next Due Date:	11-Sep-25		Operator:	SK
Equipment No.:	A-0	1-44		TE-5170		Serial No.	1316
			Ambient C	ondition			
Temperatu	re, Ta (K)	300.9	Pressure, Pa			749.2	
		Or	ifice Transfer Star	ndard Informa	ation		
Serial	No.	3864	Slope, mc	0.05914	Intercept		-0.02377
Last Calibra	ation Date:	7-Jan-25			$c = [\Delta H \times (Pa/760)]$		
Next Calibra	ation Date:	7-Jan-26	($Qstd = \{ [\Delta H \ x]$	(Pa/760) x (298/7	Γa)] ^{1/2} -bc} / m	ıc
			6				
			Calibration of	ISP Sampler		111/0	
Calibration	ΔH (orifice),		fice	Qstd (CFM)	AW (III/G) :-	HVS	60) x (298/Ta)] ^{1/2}
Point	in. of water	[ΔH x (Pa/76	$[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$		ΔW (HVS), in. of water		60) x (298/1a)] Z -axis
1	13.4		3.62	61.56	9.1		2.98
2	11.4		3.34	56.81	7.3		2.67
3	9.5		3.05	51.90	5.6		2.34
4	6.2		2.46	42.00	3.7	1.90	
5	3.6		1.87	32.10	2.1		1.43
Slope , mw = Correlation	coefficient* =	0	.9961	Intercept, bw =	-0.258	36	
*If Correlation C	Coefficient < 0.99	90, check and rec	ealibrate.				
			Set Point Ca	alculation			
	eld Calibration C sion Equation, th	ne "Y" value acco		(Pa/760) x (29	98/Ta)] ^{1/2}		
Therefore, Se	et Point; W = (m	w x Qstd + bw)	² x (760 / Pa) x (7	Га / 298) =	3.94		
Remarks:							
Conducted by:	Wong Sh	ing Kwai	Signature:	K	X -	Date:	11-Jul-25
Checked by:	Henry	Leung	Signature:	\-la	, X27	Date:	11-Jul-25

High-Volume TSP Sampler 5-POINT CALIBRATION DATA SHEET



File No. MA20003/41/0030 KTD 2D - Next to the SOR Office of Trunk Road T2 in Kai Tak Area Project No. 11-Jul-25 Next Due Date: 11-Sep-25 Operator: SK Date: Equipment No.: _____ A-01-41 Model No.: TE 5170 Serial No. 5280 **Ambient Condition** 749.2 Temperature, Ta (K) 300.9 Pressure, Pa (mmHg) **Orifice Transfer Standard Information** 0.05914 Intercept, bc 3864 Slope, mc -0.02377 Serial No. mc x Qstd + bc = $[\Delta H \times (Pa/760) \times (298/Ta)]^{1/2}$ 7-Jan-25 Last Calibration Date: Qstd = $\{ [\Delta H \times (Pa/760) \times (298/Ta)]^{1/2} -bc \} / mc$ Next Calibration Date: 7-Jan-26 **Calibration of TSP Sampler** Orfice Calibration $[\Delta W \times (Pa/760) \times (298/Ta)]^{1/2}$ ΔH (orifice), Ostd (CFM) ΔW (HVS), in. $[\Delta H \ x \ (Pa/760) \ x \ (298/Ta)]^{1/2}$ Point in. of water X - axis of water Y-axis 3.03 1 13.6 3.64 62.02 9.4 11.4 8.2 2.83 2 3.34 56.81 9.6 3.06 52.17 6.1 2.44 4 7.2 2.65 45.23 4.2 2.02 5 4.2 2.3 2.02 34.64 1.50 By Linear Regression of Y on X Slope , mw = 0.0578 Intercept, bw = -0.5358 Correlation coefficient* = *If Correlation Coefficient < 0.990, check and recalibrate. **Set Point Calculation** From the TSP Field Calibration Curve, take Qstd = 43 CFM From the Regression Equation, the "Y" value according to mw x Qstd + bw = $[\Delta W \times (Pa/760) \times (298/Ta)]^{1/2}$ Therefore, Set Point; W = $(\text{mw x Qstd} + \text{bw})^2 \times (760 / \text{Pa}) \times (\text{Ta}/298) =$ 3.89 Remarks: Signature: Date: 11-Jul-25

Signature: Date: 11-Jul-25 Conducted by: Wong Shing Kwai

Checked by: Henry Leung

CINOTECH CONSULTANTS LIMITED

Digital Dust Indicator



Date of Calibration 18-Aug-25

Certificate of Calibration

Description:

It is certified that the item under calibration has been calibrated by corresponding calibrated High Volume Sampler

Manufacturer:	Sibata Scientific Technology LTD.	Validity of Calibration Record18-Oct-25			
Model No.:	LD-5R				
Serial No.:	4X7589				
Equipment No.:	SA-01-12	Sensitivity	1 CPM = 0.001 mg	g/m3	
High Volume Sa	ampler No.: <u>A-01-03</u>	Before Sensitivity Adjustment 683 CPM			
Tisch Calibration	n Orifice No.: 3864	After Sensitiv	vity Adjustment	683 CPM	
	Ca	libration of 1	hr TSP		
Calibration	Laser Dust Monitor	•		HVS	
Point	Mass Concentration (μg/s	Mas	ss concentration (µ	ıg/m³)	
1	33.0		Y-axis 55.0		
2	44.0		71.0		
3	64.0		106.0		
Average	47.0		77.3		
Correlation co		t Correlation	_ Factor		
Particaulate Con	centration by High Volume Sampler (l'actor	77.3	
	centration by Dust Meter (µg/m³)	(F-8)	47.0		
Measureing time	e, (min)		60.0		
Set Correlation I	Factor, SCF				
SCF = [K=Hig	h Volume Sampler / Dust Meter, (μ	g/m3)]	1.6		
The Dust Monitor Factor (CF) betw	I in according to the instruction manual or was compared with a calibrated Higween the Dust Monitor and High Volumers are weighted by HOKLAS laborated	gh Volume San me Sampler.	•	was used to gener	ate the Correlation
Calibrated by:	al Officer (Wong Shing Kwai)	_	Approved by: Projec	Ct Manager (Henry	Leung)